

TEST REPORT

Issued to :

Sippyshots Private Limited
 Tashka Ka Majhra Teh Sadar,
 Pahari Gate, Rampur-244901
 Uttar Pradesh, India

J.O. No. : W20260601-043-092
 ULR Code : TC5503 26 2 00024661 F
 Issue Date : 06-06-2026
 Sample Receipt Date : 01-06-2026
 Account Manager : BD Team 4
 Credit Manager : Septesh Kumar



Customer Provided Information:#

Nature/Name of the Sample
 Location
 Source of Water

: **Packaged Drinking Water**
 : Office Premises
 : Ground Water

Laboratory Provided Information:

Sample Quantity & Packaging
 Test started on
 Test Completed
 Method of Sampling
 Date of Sampling
 Sample Picked by

: 1Litre x 20, Pet Bottle
 : 01-06-2026
 : 06-06-2026
 : FL/SOP/B/D-3
 : 01-06-2026
 : Mr. Ravikant (FARE LABS Representative) at the
 premises of Sippyshots Private Limited



Analysis Report

Physical Description : Water is clear without any sediments, suspended particles & extraneous matter.

S. No.	Parameters	Unit	Test Results	Specifications as per IS-14543:2024	Method
Microbiological Analysis					
1	Aerobic Microbial Count (20-22°C for 72 hrs.)	cfu per ml	29	NMT 100	IS: 5402 (P-1)
2	Aerobic Microbial Count (37°C for 24 hrs.)	cfu per ml	8	NMT 20	IS: 5402 (P-1)
3	Coliform	per 250ml	Absent	Absent	IS 15185
4	Faecal streptococci	per 250ml	Absent	Absent	IS:15186
5	Pseudomonas aeruginosa	per 250ml	Absent	Absent	IS 13428 (Annex D)
6	S.aureus	per 250ml	Absent	Absent	IS:5887 (P-2)
7	Salmonella sp.	per 250ml	Absent	Absent	IS:15187
8	Shigella sp.	per 250ml	Absent	Absent	IS:5887 (P-7)
9	Sulphite reducing anaerobes	per 50ml	Absent	Absent	IS:13428 (Annex C)
10	V. parahaemolyticus	per 250ml	Absent	Absent	IS:5887 (P-5) (Annex D)
11	Vibrio cholera	per 250ml	Absent	Absent	IS:5887 (P-5)
12	Yeast & Mould	per 250ml	Absent	Absent	IS 16069 (P-1)

Meenakshi Bakshi

Authorised Signatory

Ms. Meenakshi Bakshi, Scientist B

Page 1 of 3

NOTE: The laboratory accepts the responsibility for content of report. The results contained in this test report pertain only to the sample tested not for the whole lot. Test report shall not be reproduced except in full, without written approval of the laboratory. This report is intended only for your guidance, and not for legal purposes, commercial decision, and for advertisement. Samples will be destroyed after 15 days from the date of issue of test certificate unless otherwise specified. Any complaints about this report should be communicated in writing within 7 days of issue of this report. Total liability at FARELABS Pvt. Ltd. is limited to invoiced amount only. The recipient of this report is requested to kindly confirm with FARE Labs, the authenticity and genuineness of this report at report.confirmation@farelabs.com. FARE Labs is not responsible for any unconfirmed Test Report. If you have any complaint/feedback regarding the sample collection/testing/test report, please send an email at feedback@farelabs.com and call at +91 9312664533. The Sample is not drawn by FARE Labs Pvt. Ltd.

#: The details are received from Customer on its own responsibility. FARE Labs Pvt. Ltd. does not Confirm about it and hence does not take any responsibility whatsoever.

TEST REPORT

Issued to :

Sippyshots Private Limited
 Tashka Ka Majhra Teh Sadar,
 Pahari Gate, Rampur-244901
 Uttar Pradesh, India

J.O. No. : W20260601-043-092
 ULR Code : TC5503 26 2 00024661 F
 Issue Date : 06-06-2026
 Sample Receipt Date : 01-06-2026
 Account Manager : BD Team 4
 Credit Manager : Septesh Kumar



Customer Provided Information:#

Nature/Name of the Sample
 Location
 Source of Water

: **Packaged Drinking Water**
 : Office Premises
 : Ground Water

Laboratory Provided Information:

Sample Quantity & Packaging
 Test started on
 Test Completed
 Method of Sampling
 Date of Sampling
 Sample Picked by

: 1Litre x 20, Pet Bottle
 : 01-06-2026
 : 06-06-2026
 : FL/SOP/B/D-3
 : 01-06-2026
 : Mr. Ravikant (FARE LABS Representative) at the
 premises of Sippyshots Private Limited



Analysis Report

S. No.	Parameters	Unit	Test Results	Specifications as per IS-14543:2024	Method
Chemical Analysis					
1	Anionic surface active agents (as MBAS)	mg/L	BLQ, (LOQ-0.01)	0.2 Max.	APHA 24 th ED.
2	Borate (as B)	mg/L	BLQ, (LOQ-0.01)	5 Max.	IS:3025 (P-2)
3	Calcium (as Ca)	mg/L	6.4	75 Max.	APHA 24 th ED.
4	Mineral Oil	mg/L	BLQ, (LOQ-0.1)	1.0 Max.	IS:3025 (P-39)
5	Nitrate (as NO ₃)	mg/L	BLQ, (LOQ-0.1)	45 Max.	APHA 24 th ED.
6	Nitrite (as NO ₂)	mg/L	BLQ, (LOQ-0.002)	0.02 Max.	APHA 24 th ED.
7	Phenolic compounds (as C ₆ H ₅ OH)	mg/L	BLQ, (LOQ-0.0005)	0.001 Max.	APHA 24 th ED.
8	Sulphide (as H ₂ S)	mg/L	BLQ, (LOQ-0.01)	0.05 Max.	APHA 24 th ED.

Authorised Signatory
 Ms. Nayla Tazeen, Scientist-B

Page 2 of 3

NOTE: The laboratory accepts the responsibility for content of report. The results contained in this test report pertains only to the sample tested not for the whole lot. Test report shall not be reproduced except in full, without written approval of the laboratory. This report is intended only for your guidance, and not for legal purposes, commercial decision, and for advertisement. Samples will be destroyed after 15 days from the date of issue of test certificate unless otherwise specified. Any complaints about this report should be communicated in writing within 7 days of issue of this report. Total liability at FARELABS Pvt. Ltd. is limited to invoiced amount only. The recipient of this report is requested to kindly confirm with FARE Labs, the authenticity and genuineness of this report at report.confirmation@farelabs.com. FARE Labs is not responsible for any unconfirmed Test Report. If you have any complaint/feedback regarding the sample collection/testing/test report, please send an email at feedback@farelabs.com and call at +91 9312664533. The Sample is not drawn by FARE Labs Pvt. Ltd.

#: The details are received from Customer on its own responsibility. FARE Labs Pvt. Ltd. does not Confirm about it and hence does not take any responsibility whatsoever.

TEST REPORT

Issued to :

Sippyshots Private Limited
 Tashka Ka Majhra Teh Sadar,
 Pahari Gate, Rampur-244901
 Uttar Pradesh, India

J.O. No. : W20260601-043-092
 ULR Code : TC5503 26 2 00024661 F
 Issue Date : 06-06-2026
 Sample Receipt Date : 01-06-2026
 Account Manager : BD Team 4
 Credit Manager : Septesh Kumar



Customer Provided Information:#

Nature/Name of the Sample
 Location
 Source of Water

: **Packaged Drinking Water**
 : Office Premises
 : Ground Water

Laboratory Provided Information:

Sample Quantity & Packaging
 Test started on
 Test Completed
 Method of Sampling
 Date of Sampling
 Sample Picked by

: 1Litre x 20, Pet Bottle
 : 01-06-2026
 : 06-06-2026
 : FL/SOP/B/D-3
 : 01-06-2026
 : Mr. Ravikant (FARE LABS Representative) at the
 premises of Sippyshots Private Limited



Analysis Report

S. No.	Parameters	Unit	Test Results	Specifications as per IS-14543:2024	Method
Metal Analysis					
1	Aluminium (as Al)	mg/L	BLQ, (LOQ-0.01)	0.03 Max.	FL/SOP/12-01
2	Antimony (as Sb)	mg/L	BLQ, (LOQ-0.002)	0.005 Max.	FL/SOP/12-01
3	Barium (as Ba)	mg/L	BLQ, (LOQ-0.01)	0.7 Max.	FL/SOP/12-01
4	Copper (as Cu)	mg/L	BLQ, (LOQ-0.01)	0.05 Max.	FL/SOP/12-01
5	Iron (as Fe)	mg/L	0.073	0.1 Max.	FL/SOP/12-01
6	Magnesium (as Mg)	mg/L	5.81	30 Max.	FL/SOP/12-01
7	Manganese (as Mn)	mg/L	BLQ, (LOQ-0.005)	0.1 Max.	FL/SOP/12-01
8	Zinc (as Zn)	mg/L	BLQ, (LOQ-0.01)	5 Max.	FL/SOP/12-01
Pesticide Residues					
1	Pesticide residues considered individually	mg/L	BLQ, (LOQ-0.00001)	NMT 0.001	FL/SOP/03-02 & FL/SOP/05-02
2	Total Pesticide Residues	mg/L	Not Detected	0.0005 Max.	By Calculation

NMT= Not More Than

BLQ= Below Limit of Quantification; **LOQ**= Limit of Quantification

Note: FL/SOP/12-01 Based on APHA 24th Ed.

Note: FL/SOP/03-02, FL/SOP/05-02 Based on AOAC 990.06.

Opinion: Selected parameters were asked to perform and they **conforms** as per [IS-14543:2024] of Packaged Drinking water for given selected parameters only.

Authorised Signatory

Mr. Mithilesh Kumar, Scientist-B

Page 3 of 3

NOTE: The laboratory accepts the responsibility for content of report. The results contained in this test report pertains only to the sample tested not for the whole lot. Test report shall not be reproduced except in full, without written approval of the laboratory. This report is intended only for your guidance, and not for legal purposes, commercial decision, and for advertisement. Samples will be destroyed after 15 days from the date of issue of test certificate unless otherwise specified. Any complaints about this report should be communicated in writing within 7 days of issue of this report. Total liability at FARELABS Pvt. Ltd. is limited to invoiced amount only. The recipient of this report is requested to kindly confirm with FARE Labs, the authenticity and genuineness of this report at report.confirmation@farelabs.com. FARE Labs is not responsible for any unconfirmed Test Report. If you have any complaint/feedback regarding the sample collection/testing/test report, please send an email at feedback@farelabs.com and call at +91 9312664533. The Sample is not drawn by FARE Labs Pvt. Ltd.

#: The details are received from Customer on its own responsibility. FARE Labs Pvt. Ltd. does not Confirm about it and hence does not take any responsibility whatsoever.

*****End of Report*****